Ref No: SPS-CL-20201 Date: 14-JAN-2020

Author: MG, JB. Version: 2



Specification Sheet

TPP Crosslinking Agent

| Product description | A solution for crosslinking of bioinks containing chitosan, for example Chitoink. Contains 14 mM sodium tricalcium phosphate. For description on how to crosslink the desired bioink, follow the Bioprinting Protocol for that specific bioink. | | |
|------------------------|--|--|--|
| Intended use | Solution for crosslinking, <i>Research Grade</i> . For research use ON | | |
| | Not intended for in vitro diagnostics or in vivo uses. Not intended | | |
| | for administration in humans or animals. Produced under sterile | | |
| | and aseptic conditions. | | |
| Product number | CL20201 | | |
| Shelf life | 6 months, expiration date stated on package. | | |
| Storage and | Store at 4-25°C. DO NOT FREEZE. Avoid temperature | | |
| handling | fluctuations. | | |
| | Handle in accordance with good hygiene and laboratory safety | | |
| Safety | practices. Read Safety Data Sheet for more information regarding | | |
| , | ingredients and potential hazardous compounds. | | |
| Related | Safety Data Sheet can be downloaded from our website | | |
| documents | https://cellink.com/product/chitoink/ or scan the QR code below. | | |







| Property | Specification | Method |
|-----------------|-----------------------|---|
| Appearance | Transparent liquid | Visual inspection. |
| Sterility | Sterile | Tested for the presence of bacteria, fungi and yeast. |
| Endotoxin level | <10 EU/mL | Limulus Amoebocyte Lysate assay, Pharmacopoeia 2.6.14 "Bacterial endotoxins": Method D, accredited by SWEDAC. Accreditation Certification 1240: ISO 15189, 2010-11-22. |
| рН | 7.4-7.6 | Assessed with pH meter. |